



King's Research Portal

DOI:

[10.1016/j.yjmcc.2015.07.009](https://doi.org/10.1016/j.yjmcc.2015.07.009)

Document Version

Early version, also known as pre-print

[Link to publication record in King's Research Portal](#)

Citation for published version (APA):

Eykyn, T. R., Akseptijevi, D., Aughton, K. L., Southworth, R., Fuller, W., & Shattock, M. J. (2015). Multiple quantum filtered ^{23}Na NMR in the Langendorff perfused mouse heart: Ratio of Triple/Double quantum filtered signals correlate with $[\text{Na}]_i$. *Journal of Molecular and Cellular Cardiology*. 10.1016/j.yjmcc.2015.07.009

Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

Take down policy

If you believe that this document breaches copyright please contact librarypure@kcl.ac.uk providing details, and we will remove access to the work immediately and investigate your claim.

Multiple quantum filtered ^{23}Na NMR in the Langendorff perfused mouse heart: Ratio of Triple/ Double quantum filtered signals correlate with $[\text{Na}]_i$

Thomas R. Eykyn^{1,2*}, Dunja Aksentijević^{2,3*}, Karen Aughton^{2,3},
Richard Southworth^{1,2}, William Fuller⁴, Michael J. Shattock^{2,3}

¹ Division of Imaging Sciences and Biomedical Engineering, King's College London,
The Rayne Institute, St. Thomas Hospital, UK

² British Heart Foundation Centre of Research Excellence,
King's College London, UK

³ Cardiovascular Division, King's College London, UK

⁴ Division of Cardiovascular and Diabetes Medicine,
University of Dundee, Dundee, UK

* contributed equally

Correspondence: thomas.eykyn@kcl.ac.uk

Abstract

We investigate the potential of multiple quantum filtered (MQF) ^{23}Na NMR to probe intracellular $[\text{Na}]_i$ in the Langendorff perfused mouse heart. In the presence of Tm(DOTP) shift reagent the triple quantum filtered (TQF) signal originated largely from the intracellular sodium pool with a $32 \pm 6 \%$ contribution of the total TQF signal arising from extracellular sodium, whilst the rank 2 double-quantum filtered signal (DQF), acquired with a 54.7° flip-angle pulse, originated exclusively from the extracellular sodium pool. Given the different cellular origins of the ^{23}Na MQF signals we propose that the TQF/DQF ratio can be used as a semi-quantitative measure of $[\text{Na}]_i$ in the mouse heart. We demonstrate a good correlation of this ratio with $[\text{Na}]_i$ measured with shift reagent at baseline and under conditions of elevated $[\text{Na}]_i$. We compare the measurements of $[\text{Na}]_i$ using both shift reagent and TQF/DQF ratio in a cohort wild type mouse hearts and in a transgenic PLM 3SA mouse harbouring a non-phosphorylatable phospholemman, showing a modest but measurable elevation of baseline $[\text{Na}]_i$ in this model. MQF filtered ^{23}Na NMR is a potentially useful tool for studying normal and pathophysiological changes in $[\text{Na}]_i$, particularly in transgenic mouse models with altered Na regulation.

Keywords: Multiple quantum filtered ^{23}Na , TQF, DQF, Langendorff perfused mouse heart, shift reagent

Introduction

Intracellular Na concentration $[Na]_i$ is a key modulator of cardiac cell function [1]. At steady-state $[Na]_i$ has been measured between 4-16mM depending on species and preparation [2]. The Na electrochemical gradient across the cell membrane provides the energy for the action potential upstroke, as well as the active transport of many other ions, amino acids and substrates into the cell [2, 3]. $[Na]_i$ is also critical for the control of intracellular calcium $[Ca]_i$ via the sodium-calcium exchanger (NCX), thereby determining sarcoplasmic reticulum (SR) Ca content and cardiac contractility [4]. Maintenance of the Na^+ gradient is therefore fundamentally important in normal physiology but is also critically important in cardiac hypertrophy and heart failure where the elevation of $[Na]_i$ contributes to contractile and electrical dysfunction [5-7].

Historically, many techniques have been used to measure intracellular Na but few are physiologically relevant, either measuring Na at room temperature or in quiescent preparations [1]. ^{23}Na Nuclear Magnetic Resonance (NMR) is able to distinguish intra versus extracellular Na pools, employing paramagnetic shift reagents such as Tm(DOTP) [8] to separate the two. However, these reagents are efficient chelators of Ca and Mg leading to modified ion homeostasis and reduced cardiac contractility [9]. As a result, shift reagents exhibit significant toxicity precluding their use *in vivo*. In contrast, multiple quantum filtered (MQF) ^{23}Na nuclear magnetic resonance has shown great potential to probe intra and extracellular pools of Na in the *absence* of shift reagent and therefore under more physiological conditions [10, 11]. Much work has been carried out developing multiple quantum filtered ^{23}Na NMR in the perfused rat heart [10, 12, 13], *in vivo* in skeletal muscle and the brain [14] and in tumours [15]. In the perfused rat heart, the ^{23}Na TQF signal has contributions from the intra and extracellular Na pools [11, 16, 17] and has been shown to be proportional to intracellular $[Na]_i$ measured by atomic absorption spectroscopy [13]. Methods have been developed to suppress the extracellular contribution to the TQF signal [18]. On the other hand, the ^{23}Na DQF signal was found previously to arise only from the extracellular Na pool in the rat heart suggesting that TQF and DQF ^{23}Na NMR signal arise from different cellular compartments [19].

While the isolated rat heart has historically been an important model for cardiovascular research, the ability to rapidly and efficiently manipulate the genome of

the mouse has recently made this the species of choice for many physiological and biochemical studies. The application of MQF ^{23}Na NMR techniques to the isolated perfused mouse heart are much more challenging due to technical issues of the smaller heart size and, prior to this study, have not been described. The aim of the present study was therefore to revisit the utility of the multiple quantum filtered experiment and to establish and characterise the methods necessary to apply MQF ^{23}Na NMR measurements to isolated Langendorff-perfused mouse hearts in both wild-type and a transgenic PLM 3SA genotype displaying altered $[\text{Na}]_i$ homeostasis.

Methods

Langendorff mouse heart perfusion: C57BL/6J male mice (~28g body weight) were purchased from Harlan (Harlan, UK). PLM 3SA mice were derived from a colony as previously described (ref). Mice were kept in individually ventilated cages with a 12h light-dark cycle, controlled humidity and temperature (20-22°C), fed standard chow and water *ad libitum*. All experiments were approved by institutional ethical review committee and conform to the UK Home Office Guidance on the Operation of the Animals Scientific Procedures Act 1986 (HMSO).

Mice were administered terminal anaesthesia *via* intra-peritoneal pentobarbitone injection (~140 mg/kg body weight). Hearts were rapidly excised, cannulated and perfused in isovolumic Langendorff mode and perfused at 80 mmHg pressure maintained by a STH peristaltic pump controller feedback system (AD Instruments, Oxford, UK), at 37°C with Krebs-Henseleit (KH) buffer continuously gassed with 95% O_2 /5% CO_2 (pH 7.4) containing (in mM): NaCl 116, KCl 4.7, $\text{MgSO}_4 \cdot 7\text{H}_2\text{O}$ 1.2, NaHCO_3 25, KH_2PO_4 1.4, CaCl_2 1.4, glucose 11 [20]. Cardiac function was assessed using a fluid-filled cling-film balloon inserted into left ventricle (LV), and connected via a rigid polyethylene line to an MR-compatible pressure transducer (DTX Plus TNF-R, Becton-Dickinson) and a PowerLab system (AD Instruments, Australia). The volume in the intraventricular balloon was adjusted using a 1.0 ml syringe to achieve an initial LV diastolic pressure of ~9mmHg. Left ventricular developed pressure (LVDP) was calculated from the difference between systolic (SP) and diastolic pressures (DP). Functional parameters (systolic pressure, end-diastolic pressure, heart rate, LVDP,

coronary flow, perfusion pressure) were monitored using LabChart software v.7 (AD Instruments, Australia) throughout the experiment [21].

The low flow rates and the long perfusion lines in NMR systems can create problems with adequate temperature control. In addition, these long perfusion lines create significant 'dead-spaces' (where temperature, gas solubility and pH may change) and introduce long delays on switching solutions. These problems were solved in the current system by the use of two parallel continuously recirculating water-jacketed perfusion lines (Figure 1a). Control, and test solutions were gassed at 37°C before being pumped (Gilson Minipuls 3) through two parallel lines within a water-jacketed (37°C) umbilical line (approx 3 metres in length). These two lines then enter a novel three-way pneumatically operated valve positioned directly above the cannula (Figure 1a). Control solution was passed via the pneumatic valve to the heart while the test solution was recirculated back up the umbilicus to the external reservoir. When the valve is activated, the control solution is recirculated and the test solution passes to the heart with effectively no distal stagnant dead-space and a switching time of <2sec, at the level of the heart. In order to protect the magnet, coils etc from accidental leakage, the perfusion system includes a 'flood plain' between the perfusion lines, valve etc and the cannula (Fig. 1a). This flood plain includes a moisture sensor connected to a remote alarm. The umbilical line, transducers, valve and all fittings were manufactured exclusively of non-magnetic parts and was fed *via* the top of the magnet and lowered into the active region of the bore/microimaging coil. At the end of each experiment, the perfusion system was removed from the magnet and hearts were immediately snap frozen using Wollenberger tongs, pre-cooled in liquid nitrogen, and wet and dry weights were recorded.

NMR Spectroscopy: All experiments were carried out on a Bruker Avance III 400 MHz wide-bore spectrometer (Bruker, Karlsruhe, Germany) equipped with triple-axis gradients, a microimaging probe and exchangeable RF coil inserts (10mm ^{23}Na coil or $^1\text{H}/^{31}\text{P}$ dual tune coil) that enable rapid switching between different nuclei. NMR tubes (O.D. 10mm) (Wilma, UK) were shortened in length to just greater than the NMR active coil region in order to minimize the perfusate dead volume beneath the heart. The bore of the NMR was maintained (~313K) by continuous delivery of warm water,

from a thermostatically controlled water bath, through the imaging gradients of the MR. Temperature calibration was performed *in situ* using a capillary containing ethylene glycol at the position of the heart.

To analyse the relationship between $[Na]_i$ elevation and TQF and DQF signal intensities, hearts were equilibrated for 20mins, followed by 20min treatment with: (i) standard KHB (n=6) (ii) 50 μ M ouabain (n=6), (iii) K^+ -free buffer (n=3), (iv) K^+ -free/ Ca^{2+} -free (n=3) or (v) K^+ -free/ Ca^{2+} -free/ Mg^{2+} -free (n=3) buffer. Interleaved TQF and DQF NMR acquisitions were recorded throughout the stability period and during the intracellular sodium elevation protocols. At the end of each intervention, perfusate was switched to KHB containing 5mM Tm(DOTP)⁵⁻ to shift the extracellular $[Na]_e$ signal and quantify $[Na]_i$. An intracellular volume of 2.5 ml/g dry weight of tissue was assumed for calculation of $[Na]_i$ [22].

In order to confirm that cardiac energetics were normal and not compromised prior to any elevation in $[Na]_i$, fully relaxed ³¹P experiments were also performed with the same perfusion system and phosphate free buffer, acquired with a 60° flip angle, 256 scans, a repetition time of 3.8s and a total experiment duration of 16mins.

Multiple Quantum Filtering ²³Na MRS: The multiple quantum filtered pulse sequence consists of [10]: $90^\circ(\phi_1) - \tau_m/2 - 180^\circ(\phi_2) - \tau_m/2 - \beta^\circ(\phi_2) - \delta - \beta^\circ(\phi_3) - FID(\phi_{rec})$. DQF experiments were acquired with a four-step phase cycle: $\phi_1 = \phi_2 = 90^\circ, 180^\circ, 270^\circ, 0^\circ$; $\phi_3 = 0^\circ$; $\phi_{rec} = 0^\circ, 180^\circ, 0^\circ, 180^\circ$. The flip angle β was set either to the magic angle, 54.7° (rank 2 $T_{2,\pm 2}$ only) or to 90° (both rank 2 $T_{2,\pm 2}$ and rank 3 $T_{3,\pm 2}$). TQF experiments were acquired with a six-step phase cycle: $\phi_1 = 30^\circ, 90^\circ, 150^\circ, 210^\circ, 270^\circ, 330^\circ$; $\phi_2 = 120^\circ, 180^\circ, 240^\circ, 300^\circ, 0^\circ, 60^\circ$; $\phi_3 = 0^\circ$; $\phi_{rec} = 0^\circ, 180^\circ, 0^\circ, 180^\circ, 0^\circ, 180^\circ$ with the flip angle $\beta = 90^\circ$ (rank 3 $T_{3,\pm 3}$ only). The mixing time ($\tau = 3.6ms$) was calibrated for the maximum TQF signal in the mouse heart and set to be the same for the DQF experiments. All MQF experiments were acquired with 192 scans, 2048 data points, sweep width of 50ppm, an acquisition time of 200ms, pre-scan delay of 200ms and a total acquisition time of 1.24 mins. An exponential line broadening factor of 10Hz was applied prior to Fourier transformation and subsequent baseline correction.

Shift reagent ^{23}Na MRS : All hearts were perfused at the end of the protocol with modified KHB containing 5mM Tm(DOTP) $^{5-}$ (thulium (III) 1,4,7,10-tetraazacyclododecane-N,N',N'',N'''tetra(methyl-enephosphonate)) (Macrocyclics, Dallas, USA). ^{23}Na acquisitions were recorded in real-time during the perfusion with the shift reagent until the extracellular $[\text{Na}]_e$ peak was completely shifted (5-10mins) (Figure 2b). Following infusion of shift reagent a single quantum spectrum was acquired with 128 scans, 2048 data points, sweep width of 50ppm, an acquisition time of 200ms, pre-scan delay of 200ms and a total acquisition time of 50s. An exponential line broadening factor of 10Hz was applied prior to Fourier transformation and subsequent baseline correction. The resulting $[\text{Na}]_i$ peak was referenced to a glass bulb containing 2.9 μmol NaCl with 10mM Tm(DOTP) $^{5-}$ at the position of the heart. The Na present in the calibration bulb was assumed to be 100% NMR visible ie $f_o = 1.0$. The visibility of intracellular Na is more controversial but, in these studies, was assumed to be $f_i = 0.4$ and this was assumed to remain constant throughout the experimental protocol. [23-25].

Statistical Analysis: Data were analysed and expressed as mean \pm standard error of the mean (SEM), n = number of independent experiments. Where appropriate, statistical analysis was performed using Students t-test and relationship between variables analysed using Pearson correlation. $P < 0.05$ was considered statistically significant.

Results

Table 1 shows functional data at 5 min intervals during the 20 min stability period. Averaged over the 20 min period (\pm SEM), systolic pressure was 107 ± 6 mmHg, end diastolic pressure (EDP) was 11 ± 2 mmHg, left ventricular developed pressure (LVDP) was 96 ± 15 mmHg, heart rate (HR) was 455 ± 66 bpm and coronary flow was 2.3 ± 0.2 ml/min.

Figure 2(a) shows a representative ^{31}P spectrum acquired in the mouse heart at baseline prior to intervention and is representative of normal cardiac energetics displaying resonances from adenosine triphosphate (ATP), phosphocreatine (PCr) and inorganic phosphate (Pi). Figure 2(b) shows a time-course of ^{23}Na spectra acquired during perfusion with Tm(DOTP) $^{5-}$. In the absence of shift reagent (time=0) the total Na

peak corresponds to the sum of contributions from intracellular $[Na]_i$, extracellular/intravascular $[Na]_e$, and from buffer surrounding the heart $[Na]_o$. During infusion of shift reagent the ^{23}Na signal from the extracellular/intravascular $[Na]_e$ pool is shifted first and at later times (5-10 mins) mixing of buffer surrounding the heart is achieved.

Figure 3(a) shows the single quantum ^{23}Na NMR spectrum acquired after infusion of shift reagent where an unshifted Na signal arising from the intracellular $[Na]_i$ pool, can readily be discerned and quantified. Fig. 3(b-d) shows representative ^{23}Na MQF spectra acquired in the presence of shift reagent where the positions of the relevant peaks are denoted by dashed lines. Fig. 3(b) shows the TQF signal arising principally from the intracellular $[Na]_i$ pool with a smaller contribution from the extracellular pool. The contribution to the total TQF signal arising from Na_e acquired with shift reagent was $Na_e/(Na_e+Na_i) = 0.32 \pm 0.06$ (S.E. $n=6$) corresponding to 32% of the total TQF signal. That is, in these experiments, 68% of the TQF signal reflects intracellular Na while 32% of the signal comes from extracellular sources. Fig. 3(c) shows the DQF spectrum acquired with shift reagent and a flip angle $\beta = 54.7^\circ$ showing exclusively extracellular Na_e . Intracellular sodium was not observed employing the DQF sequence with a magic flip angle pulse. Fig. 3(d) shows the DQF spectrum acquired with shift reagent and a flip angle $\beta = 90^\circ$ showing both intracellular and extracellular Na and was not used further in this work due to a mixed contribution from both intra and extracellular Na pools.

Figure 4 (a) shows the amplitude of the TQF and DQF signals acquired in an interleaved fashion during 20 mins stabilisation period and during 20 mins infusion of $50\mu M$ ouabain. Representative ^{23}Na TQF and DQF spectra are shown in Fig. 4(b) during baseline stability, (c) after 20 mins infusion of ouabain (d) after 20 mins K^+ -free buffer and (e) after 20 mins K^+ -free/ Ca^{2+} -free buffer. A significant increase in the TQF signal was observed with respect to the baseline TQF signal with minimal changes in the DQF signal. The increase in $[Na]_i$ is consistent with the mode of action of ouabain and blocking of the Na/K ATPase or inhibition of Na efflux using modified buffers.

Figure 5 shows a plot of the ratio of the TQF/DQF signals ($R^{TQF/DQF}$) as a function of $[Na]_i$ measured with shift reagent, in control hearts at baseline $[Na]_i$ and in the four groups displaying elevated $[Na]_i$ after treatment with $50\mu M$ ouabain ($n=6$); K^+ -free ($n=3$), K^+ -free/ Ca^{2+} -free ($n=3$) and K^+ -free/ Ca^{2+} -free/ Mg^{2+} -free ($n=3$). The mean $R^{TQF/DQF}$

was 3.0 ± 0.3 for control hearts, 4.4 ± 0.7 for 50 μ M ouabain treatment, 5.2 ± 1.1 for K⁺-free hearts, 7.7 ± 1.1 for K⁺-free/Ca²⁺-free and 6.3 ± 0.4 for K⁺-free/Ca²⁺-free/Mg²⁺-free with a measured [Na]_i = 15.7 ± 1.6 mM; 25.5 ± 5.8 mM; 40.0 ± 5.7 mM, 57.9 ± 2.4 mM and 54.9 ± 6.5 mM, respectively. $R^{TQF/DQF}$ showed a good correlation with [Na]_i. The intercept on the y-axis demonstrates a non-zero $R^{TQF/DQF}$ at [Na]_i = 0mM confirming that there is a non-zero contribution from [Na]_e to the TQF signal. From this line of fit, intracellular Na can be estimated from a measured $R^{TQF/DQF}$ according to the equation: $[Na]_i = (R^{TQF/DQF} - 1.42)/0.10$. This suggests that the extracellular contribution to the TQF signal is ~47%, therefore slightly greater than the 32% contribution estimated earlier in the presence of shift reagent, but lower than previously reported values for the rat heart of ~64% [13].

Figure 6 shows intracellular [Na]_i measured in a cohort of wild type mice and in the PLM 3SA transgenic model. Fig. 6(a) shows [Na]_i measured with the shift reagent technique for the two groups, with a measured [Na]_i = 15.7 ± 1.6 mM (n=6) in the WT and [Na]_i = 20.5 ± 1.4 mM (n=6) in the PLM 3SA (p=0.0497). Fig. 6(b) shows [Na]_i measured from the TQF/DQF ratio using the calibration graph in Fig. 5. The TQF/DQF ratio was $R^{TQF/DQF} = 2.59 \pm 0.09$ in the WT group with an extrapolated value of [Na]_i = 11.6 ± 0.9 mM (n=37) and $R^{TQF/DQF} = 3.12 \pm 0.18$ in the PLM 3SA group with an extrapolated [Na]_i = 17.0 ± 1.8 mM (n=21) (p=0.024). Employing the TQF/DQF ratio required a larger pooled number of hearts to show a significant difference between the groups. A modest but significant elevation of baseline [Na]_i was measured in the PLM 3SA mouse compared to WT using both techniques.

Discussion

We have developed an MR-compatible Langendorff constant-pressure/constant flow perfusion system to non-invasively probe mouse myocardial [Na]_i at physiological heart rates and temperatures. Multiple quantum filtering techniques (DQF and TQF) were used to monitor [Na]_i during baseline physiological perfusion and during relative changes in [Na]_i initiated by inhibition of the Na/K ATPase activity by 50 μ M ouabain and modified KH buffers (K⁺-free, K⁺-free/Ca²⁺-free and K⁺-free/Ca²⁺-free/Mg²⁺-free KHB). Using the commonly used shift reagent, Tm(DOTP)⁵⁻ to distinguish the relative contributions to the MQF signal from intra [Na]_i and extracellular [Na]_e, we were able

to quantify $[\text{Na}]_i$ and thereby compare these to the measurements obtained using the MQF experiments. In the presence of shift reagent we found that the proportion of the total TQF signal arising from the extracellular $[\text{Na}]_e$ pool was approximately 32%, indicating that the signal mainly represents $[\text{Na}]_i$. We observed that the DQF signal was independent of elevations in $[\text{Na}]_i$, consistent with this signal arising from the extracellular $[\text{Na}]_e$ pool. In both TQF and DQF multiple quantum filtered experiments the bulk isotropic Na signal is filtered out of the spectrum leaving only the Na pools subject to anisotropic interactions; the TQF arising due to biexponential relaxation and macromolecular Na binding in the intracellular matrix whilst the rank 2 DQF signal arises due to a residual quadrupolar coupling due to association with ordered structures such as tissue fibre orientation. The relative amplitudes of the TQF and DQF signals, with the former being greater than the latter, represent the different anisotropic environments found in these different compartments. We found a good correlation between the TQF/DQF ratio $R^{\text{TQF/DQF}}$ and $[\text{Na}]_i$ and propose, therefore, that this ratio can be used as a semi-quantitative measure of $[\text{Na}]_i$ in the mouse heart. Using the DQF signal as an internal extracellular reference on a heart-by-heart basis renders the TQF/DQF ratio independent of heart weight (data not shown), does not require an external $[\text{Na}]_{\text{ref}}$ standard for calibration and enables a direct measure of $[\text{Na}]_i$ without resorting to the use of shift reagents.

Absolute quantification of $[\text{Na}]_i$ is challenging, particularly in the perfused mouse heart under physiological perfusion conditions. For the calculation of intracellular volume we assume a value of 2.5 ml/g dry weight of tissue which is adequate for the present control group of hearts [22]. This value is also in close agreement with that of 0.44ml/g wet weight estimated by Jelicks and Gupta [26]. However, exact measurement of intracellular volume would be required if these techniques are to be applied under conditions where this might change. It must also be noted that a large proportion of the intracellular pool may not be visible due to the quadrupolar nature of the intracellular $^{23}\text{Na}^+$ bound to macromolecules. The NMR invisibility of a fraction of the intracellular Na has been reported by some [27] while others have reported that intracellular Na is 100% visible [26].

While the extent of ^{23}Na visibility may bias the absolute quantification of $[\text{Na}]_i$ using shift reagent [27], multiple quantum filtered ^{23}Na will be useful to examine

relative changes in $[Na]_i$ applied to perfused mouse hearts. Both shift reagent and the TQF/DQF ratio were able to distinguish similar baseline differences in $[Na]_i$ in the two groups of hearts. The PLM 3SA genotype displayed a modest but significantly elevated $[Na]_i$ as measured both by shift reagent and the TQF/DQF ratio. However the poorer signal-to-noise ratio of the TQF and DQF NMR measurements compared to shift reagent, see Figure 3, resulted in a greater heart-to-heart variability, and therefore a larger pooled n number was required to reach significance between the two groups. There is also a slight discrepancy in the absolute magnitude of the measured $[Na]_i$ using the two techniques which may reflect the different measurement techniques as well as possible errors introduced in the derivation of the calibration curve. The difference might also reflect the different perfusion conditions, the groups measured with shift reagent being non-contracting due to the Ca chelation effect of the shift reagent whilst the TQF/DQF measurements are performed under more physiological conditions as reflected by the cardiac function in table 1. Whilst there was significant heart-to-heart variability in the TQF/DQF measurements requiring a large n number to distinguish baseline differences in $[Na]_i$, the methods developed and validated in our study do not require the use of shift reagent, thus are performed under more physiological conditions, making our study readily applicable to the study of $[Na]_i$ regulation in a wealth of transgenic models and muscle types with chronically altered $[Na]_i$. If absolute quantification of $[Na]_i$ is not required but relative changes with respect to baseline then the TQF measurement in the absence of DQF would also offer a viable and more reproducible alternative to the TQF/DQF ratio proposed here.

Acknowledgements

The authors acknowledge financial support from the Department of Health via the National Institute for Health Research (NIHR) comprehensive Biomedical Research Centre award to Guy's & St Thomas' NHS Foundation Trust in partnership with King's College London and King's College Hospital NHS Foundation Trust. The work was supported by the British Heart Foundation Centre of Research Excellence (RE/08/003), British Heart Foundation Programme Grant (RG/12/4/29426), the Wellcome Trust and EPSRC Centre of Excellence in Medical Engineering Centre (WT088641/Z/09/Z), and the King's College London and UCL Comprehensive Cancer Imaging Centre. The views

expressed are those of the author and not necessarily those of the NHS, the NIHR or the Department of Health. TRE is grateful for support from the CRUK and EPSRC Cancer Imaging Centre in association with the MRC and the Department of Health (England) grant C1060/A10334.

The perfusion system described here was built by the engineers in the medical physics workshop (St Thomas' Hospital) and the expertise of Mr Michael Kelly is particularly acknowledged.

Table 1: Cardiac function acquired at 5min intervals during the 20min stability period.

Functional Parameters	Time (min)				
(n=7)	0	5	10	15	20
Systolic Pressure (mmHg)	109±12	113±11	103±17	103±19	106±19
EDP(mmHg)	11±1	9±2	11±2	10±2	12±2
LVDP (mmHg)	98±12	104±11	92±16	93±18	93±19
Heart Rate (bpm)	382±32	526±54	453±67	449±79	464±96
Coronary Flow (ml/min)	2.2±0.1	2.2±0.2	2.4±0.2	2.2±0.2	2.6±0.2

Figure Captions

FIG. 1. NMR compatible Langendorff constant pressure perfusion system (a, b) and a schematic of the pneumatically driven three-way valve to allow switching between reservoirs with minimum dead-volume (c).

Fig. 2. (a) Example ^{31}P spectrum acquired during baseline stability period showing normal cardiac energetics. (b) Time-series of ^{23}Na NMR spectra acquired in a perfused mouse heart during the infusion of Tm(DOTP) shift reagent. An initial shift of the intravascular extra cellular signal $[\text{Na}]_e$ is observed followed by the buffer surrounding the heart was achieved within a 5-10 mins timeframe.

Fig. 3. (a) ^{23}Na single quantum spectrum post infusion of shift reagent where the intracellular peak from $[\text{Na}]_i$ is clearly discerned. (b) TQF spectrum acquired from the same heart as in (a) in the presence of shift reagent showing predominantly intracellular $[\text{Na}]_i$ with a smaller contribution from extracellular $[\text{Na}]_e$. (c) ^{23}Na DQF spectrum acquired with a flip angle $\beta = 54.7^\circ$ showing predominantly extracellular $[\text{Na}]_e$. (d) ^{23}Na DQF spectrum acquired with a flip angle $\beta = 90^\circ$ showing both intracellular and extracellular Na.

Fig. 4. (a) Peak areas shown for a time-series (\pm S.E. $n=6$) of ^{23}Na TQF and DQF spectra acquired in an interleaved fashion in the perfused mouse heart during stability and during subsequent infusion of $50\mu\text{M}$ ouabain, shaded area. Representative TQF and DQF spectra are shown; (b) during stability, (c) after 20 mins infusion of ouabain (d) after 20 mins K^+ -free buffer and (e) after 20 mins K^+ -free/ Ca^{2+} -free buffer.

Fig. 5. Plot of the TQF/DQF ratio $R^{\text{TQF/DQF}}$ vs the measured concentration $[\text{Na}]_i$ in the perfused mouse heart at baseline (filled square, $n=6$) and following a 20 min intervention to elevate $[\text{Na}]_i$; $50\mu\text{M}$ ouabain (filled triangle, $n=6$); 0mM K^+ (open diamond, $n=3$), 0mM K^+ / 0mM Ca^{2+} (filled circle, $n=3$) and 0mM K^+ / 0mM Ca^{2+} / 0mM Mg^{2+} (open circle, $n=3$). Mean data are shown for each group \pm S.E.

Fig. 6. Plot of measured $[\text{Na}]_i$ in the WT mouse and PLM 3SA transgenic mouse using (a) the shift reagent technique and (b) the derived values of $[\text{Na}]_i$ using the TQF/DQF ratio. Significant differences between the two groups are shown $**p<0.05$.




Figure 1




Figure 2




Figure 3




Figure 4




Figure 5




Figure 6

References

- [1] Murphy E, Eisner DA. Regulation of Intracellular and Mitochondrial Sodium in Health and Disease. *Circ Res*. 2009;104:292-303.
- [2] Despa S, Bers DM. Na⁺ transport in the normal and failing heart - Remember the balance. *J Mol Cell Cardiol*. 2013;61:2-10.
- [3] Bay J, Kohlhaas M, Maack C. Intracellular Na⁺ and cardiac metabolism. *J Mol Cell Cardiol*. 2013;61:20-7.
- [4] Ottolia M, Torres N, Bridge JHB, Philipson KD, Goldhaber JL. Na/Ca exchange and contraction of the heart. *J Mol Cell Cardiol*. 2013;61:28-33.
- [5] Kohlhaas M, Liu T, Knopp A, Zeller T, Ong MF, Bohm M, et al. Elevated cytosolic Na⁺ increases mitochondrial formation of reactive oxygen species in failing cardiac myocytes. *Circulation*. 2010;121:1606-13.
- [6] Boguslavskyi A, Pavlovic D, Aughton K, Clark JE, Howie J, Fuller W, et al. Cardiac hypertrophy in mice expressing unphosphorylatable phospholemman. *Cardiovasc Res*. 2014;104:72-82.
- [7] Despa S, Islam MA, Weber CR, Pogwizd SM, Bers DM. Intracellular Na⁺ concentration is elevated in heart failure but Na/K pump function is unchanged. *Circulation*. 2002;105:2543-8.
- [8] Weidensteiner C, Horn M, Fekete E, Neubauer S, von Kienlin M. Imaging of intracellular sodium with shift reagent aided Na-23 CSI in isolated rat hearts. *Magn Reson Med*. 2002;48:89-96.

- [9] Simor T, Lorand T, Szollosy A, Gaszner B, Digerness SB, Elgavish GA. Na-23 NMR shift reagents enhance cardiac staircase effect in isolated perfused rat hearts. *NMR Biomed.* 1999;12:267-74.
- [10] Payne GS, Seymour AM, Styles P, Radda GK. Multiple quantum filtered ²³Na NMR spectroscopy in the perfused heart. *NMR Biomed.* 1990;3:139-46.
- [11] Dizon JM, Tauskela JS, Wise D, Burkhoff D, Cannon PJ, Katz J. Evaluation of triple-quantum-filtered ²³Na NMR in monitoring of Intracellular Na content in the perfused rat heart: comparison of intra- and extracellular transverse relaxation and spectral amplitudes. *Magn Reson Med.* 1996;35:336-45.
- [12] Tauskela JS, Dizon JM, Whang J, Katz J. Evaluation of multiple-quantum-filtered ²³Na NMR in monitoring intracellular Na content in the isolated perfused rat heart in the absence of a chemical-shift reagent. *J Magn Reson.* 1997;127:115-27.
- [13] Schepkin VD, Choy IO, Budinger TF, Obayashi DY, Taylor SE, DeCampi WM, et al. Sodium TQF NMR and intracellular sodium in isolated crystalloid perfused rat heart. *Magn Reson Med.* 1998;39:557-63.
- [14] Reddy R, Bolinger L, Shinnar M, Noyszewski E, Leigh JS. Detection of residual quadrupolar interaction in human skeletal muscle and brain in vivo via multiple quantum filtered sodium NMR spectra. *Magn Reson Med.* 1995;33:134-9.
- [15] Babsky AM, Zhang H, Hekmatyar SK, Hutchins GD, Bansal N. Monitoring chemotherapeutic response in RIF-1 tumors by single-quantum and triple-quantum-filtered (²³Na MRI, (¹H diffusion-weighted MRI and PET imaging. *Magn Reson Imaging.* 2007;25:1015-23.

- [16] Tauskela JS, Dizon JM, Whang J, Katz J. Evaluation of multiple-quantum-filtered Na-23 NMR in monitoring intracellular Na content in the isolated perfused rat heart in the absence of a chemical-shift reagent. *J Magn Reson.* 1997;127:115-27.
- [17] Jelicks LA, Gupta RK. On the Extracellular Contribution to Multiple Quantum Filtered Na-23 Nmr of Perfused Rat-Heart. *Magn Reson Med.* 1993;29:130-3.
- [18] Navon G. Complete Elimination of the Extracellular Na-23 Nmr Signal in Triple Quantum Filtered Spectra of Rat Hearts in the Presence of Shift-Reagents. *Magn Reson Med.* 1993;30:503-6.
- [19] Tauskela JS, Dizon JM, Cannon PJ, Katz J. Detection of an Extracellular Contribution from a 2nd-Rank Tensor to the Double-Quantum-Filtered Na-23 Nmr-Spectrum in the Isolated-Perfused Rat-Heart. *Journal of Magnetic Resonance Series B.* 1995;108:165-9.
- [20] Sutherland FJ, Shattock MJ, Baker KE, Hearse DJ. Mouse isolated perfused heart: characteristics and cautions. *Clin Exp Pharmacol Physiol.* 2003;30:867-78.
- [21] Sutherland FJ, Shattock MJ, Baker KE, Hearse DJ. Mouse isolated perfused heart: characteristics and cautions. *Clin Exp Pharmacol Physiol.* 2003;30:867-78.
- [22] Askenasy N, Navon G. Continuous monitoring of intracellular volumes in isolated rat hearts during normothermic perfusion and ischemia. *J Magn Reson.* 1997;124:42-50.
- [23] Ogino T, den Hollander JA, Shulman RG. ³⁹K, ²³Na, and ³¹P NMR studies of ion transport in *Saccharomyces cerevisiae*. *Proc Natl Acad Sci U S A.* 1983;80:5185-9.
- [24] Castle AM, Macnab RM, Shulman RG. Measurement of intracellular sodium concentration and sodium transport in *Escherichia coli* by ²³Na nuclear magnetic resonance. *J Biol Chem.* 1986;261:3288-94.

- [25] Steenbergen C, Perlman ME, London RE, Murphy E. Mechanism of Preconditioning - Ionic Alterations. *Circ Res.* 1993;72:112-25.
- [26] Jelicks LA, Gupta RK. Multinuclear Nmr-Studies of the Langendorff Perfused Rat-Heart. *J Biol Chem.* 1989;264:15230-5.
- [27] Imahashi K, London RE, Steenbergen C, Murphy E. Male-female differences in intracellular Na⁺ regulation during ischemia/reperfusion in mouse heart. *J Mol Cell Cardiol.* 2004;37:747-53.